

Transmitter Instructions

Pre-Programmed Systems

Please make sure the transmitter inside the neck of your sensor pad is turned on prior to using product.

1. Open neck of pad and pull out transmitter.
2. Turn on the transmitter by moving the switch to the "on" position.
3. Once you have the switch in the on position, you should see the red light turn on and stay lit for a short time on the top side on the transmitter.
4. Once the red light turns off, it should blink every 45 seconds to show that it has power.
5. Now that you have the transmitter turned on, you can place it back inside the neck of the pad and seal closed with provided sticker. Please activate your pad following the steps below.

Products should be pre-programmed and ready for activation.


**Your system has been pre-programmed.
Please activate your pad prior to first use.**

To activate:

Press and hold pad for 30 seconds then release. This pad is now activated and ready for use. This is only necessary for the first-time use.

-See monitor instructions for programming

Manufactured for:



Press and hold for
30 seconds



AliMed.com
7276REV1114 CCO #729
©2014 AliMed® inc. All rights reserved.
Made in China

Transmitter Instructions

Pre-Programmed Systems

Please make sure the transmitter inside the neck of your sensor pad is turned on prior to using product.

1. Open neck of pad and pull out transmitter.
2. Turn on the transmitter by moving the switch to the "on" position.
3. Once you have the switch in the on position, you should see the red light turn on and stay lit for a short time on the top side on the transmitter.
4. Once the red light turns off, it should blink every 45 seconds to show that it has power.
5. Now that you have the transmitter turned on, you can place it back inside the neck of the pad and seal closed with provided sticker. Please activate your pad following the steps below.

Products should be pre-programmed and ready for activation.


**Your system has been pre-programmed.
Please activate your pad prior to first use.**

To activate:

Press and hold pad for 30 seconds then release. This pad is now activated and ready for use. This is only necessary for the first-time use.

-See monitor instructions for programming

Manufactured for:



Press and hold for
30 seconds



AliMed.com
7276REV1114 CCO #729
©2014 AliMed® inc. All rights reserved.
Made in China